

Date : 04/11/2023

Illumina, Inc.  
5200 Illumina Way  
San Diego CA 92122  
USA

Attn: Kelly Byron

**Re. : CU US + Canada Certificate**

Type of Equipment : Genetic Sequencer  
Model Designation : See Certificate  
Certificate No. : CU 72300009 0001  
File No. : 31181842 052  
Engineer/Contact : Gregory Chase  
Standard(s) : UL 61010-1:2012 R7.19  
UL 61010-2-081:2019  
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) +  
A1  
CSA C22.2 No. 61010-2-081:19

Dear Ms. Byron,

The above referenced technical equipment has been tested and was found to be in compliance with the listed test requirement(s). Enclosed, please find the TUV Rheinland approval document No. CU 72300009 0001. It authorizes you to label the listed product(s) with the TUV Rheinland Mark identified in the approval document. For compliance, the Test Mark must be on the approved unit.

Your product is subject to regular factory follow-up inspections as well as annual certificate and factory registration fees.

In using the TUV Rheinland Mark you are obligated to comply with the TUV Rheinland of North America Service Agreement.

If we can be of any further assistance to you, please do not hesitate to contact us.

Sincerely yours,  
Certification Body

Dipl.-Ing. M. Glagla  
QA Certification Officer

Enclosure

TUV Rheinland  
of North America, Inc.

295 Foster St.  
Suite 100  
Littleton, MA 01460

Tel 978-266-9500  
Fax 978-266-9992  
Toll Free TUV-RHEINLAND  
Mail info@tuv.com  
Web www.tuv.com

Member of  
TUV Rheinland Group

# Certificate



Certificate no.

CU 72300009 01

**License Holder:**

Illumina, Inc.  
5200 Illumina Way  
San Diego CA 92122  
USA

**Manufacturing Plant:**

Illumina, Inc.  
25861 Industrial Blvd.  
Hayward CA 94545  
USA

**Test report no.:** USA-GC 31181842 052

**Client Reference:** Kelly Byron

**Tested to:**

UL 61010-1:2012 R7.19  
UL 61010-2-081:2019  
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) + A1  
CSA C22.2 No. 61010-2-081:19

**Certified Product:** Genetic Sequencer

**License Fee - Units**

Model Designation: MiSeq, MiSeqFGx

Rated Voltage: AC 100-240V, 50/60Hz

Rated Power: 400W

Protection Class: I

Special Remarks: Replaces Certificate CU72171864.

Appendix: 1, 1-9

**Licensed Test mark:**



**Date of Issue**

**(day/mo/yr)**

11/04/2023